



National Accreditation Board for  
Testing and Calibration Laboratories

**CERTIFICATE OF ACCREDITATION**

**DTPS COAL TESTING LABORATORY, ADANI  
ELECTRICITY MUMBAI LIMITED**

has been assessed and accredited in accordance with the standard

**ISO/IEC 17025:2017**

**"General Requirements for the Competence of Testing &  
Calibration Laboratories"**

for its facilities at

4TH FLOOR, SERVICE BUILDING, DAHANU THERMAL POWER STATION, DAHANU, MAHARASHTRA,  
INDIA

in the field of

**TESTING**

Certificate Number: TC-6307

Issue Date: 15/11/2023

Valid Until:

14/11/2025

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the relevant requirements of NABL.

(To see the scope of accreditation of this laboratory, you may also visit NABL website [www.nabl-india.org](http://www.nabl-india.org))

Name of Legal Entity: Adani Electricity Mumbai Limited

Signed for and on behalf of NABL



N. Venkateswaran  
Chief Executive Officer



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

DTPS COAL TESTING LABORATORY, ADANI ELECTRICITY MUMBAI LIMITED, 4TH FLOOR,  
SERVICE BUILDING, DAHANU THERMAL POWER STATION, DAHANU, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

TC-6307

**Page No**

1 of 1

**Validity**

15/11/2023 to 14/11/2025

**Last Amended on**

23/11/2023

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Permanent Facility				
1	CHEMICAL- BUILDING MATERIAL	Fly ash	Fineness by Wet Sieving	IS 1727
2	CHEMICAL- BUILDING MATERIAL	Fly ash	Loss on Ignition	IS 1727
3	CHEMICAL- SOLID FUELS	Coal	Ash	IS 1350 (Part 1)
4	CHEMICAL- SOLID FUELS	Coal	Fixed Carbon	IS 1350 (Part 1)
5	CHEMICAL- SOLID FUELS	Coal	Gross Calorific Value	IS 1350 (Part 2)
6	CHEMICAL- SOLID FUELS	Coal	Moisture in Air Dried Sample	IS 1350 (Part 1)
7	CHEMICAL- SOLID FUELS	Coal	Sulphur	IS 1350 (Part 3)
8	CHEMICAL- SOLID FUELS	Coal	Total Moisture	IS 1350 (Part 1)
9	CHEMICAL- SOLID FUELS	Coal	Volatile Matter	IS 1350 (Part 1)